



Nothing else measures up!

Follow us!



Prospect House
George Cayley Drive
Clifton Moor
York
England
YO30 4XE

Tel: +44 (0) 1904 692723
Fax: +44 (0) 1904 690385

E-Mail: sales@yorksurvey.co.uk



Thermamite Digital Probe Thermometer



Code: 32072

Operating Instructions

Thermamite Digital Probe Thermometer

Instrument Operation:

The instrument is switched on by using the switch located on the front of the instrument. Within one second the thermometer is ready to work. Apply the tip of the probe to the substance, surface or medium to be measured. The instrument reading may take a few seconds to stabilise, depending on the nature of the measurement.

Battery Replacement:

The instrument will display 'LO BAT' to warn that the batteries will need changing soon. The instrument continues to measure accurately, but after usage the instrument will display 'flat bat' and shutdown. Replace the batteries with three AAA batteries located under the rear cover that is held in place by a single screw.

Error Message:

If the thermometer or probe has developed a fault the instrument will show 'Err'.

EMC/RFI:

Instrument performance may be affected if operated within a high frequency radio field, such as near a mobile phone, or if subjected to an electrostatic shock.

Guarantee:

This instrument carries a two year warranty against defects in either components or workmanship. During this period, products that prove to be defective will, at the manufacturer's discretion, be either repaired or replaced without charge. This warranty does not apply to probes, where a six month period is offered. The product warranty does not cover damage caused by fair wear and tear, abnormal storage conditions, incorrect use, accidental misuse, abuse, neglect, misapplication or modification. In line with our policy of continuous development, we reserve the right to amend our product specification without prior notice.

